
Midmark M11 - v1.0.0
Total Cycles: 4490

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2024 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 84.2 F 0.1
4:00 139.3 F 0.7
6:00 121.8 F 3.0
8:00 146.6 F 5.5
10:00 170.1 F 8.5
12:00 196.8 F 12.3
14:00 208.3 F 3.5
16:00 221.2 F 6.6
18:00 232.1 F 10.1
20:00 236.5 F 9.4
22:00 236.7 F 10.1
24:00 239.3 F 8.6
26:00 246.0 F 13.3
28:00 257.3 F 19.1
30:00 267.2 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.2 F 29.0
2:00 272.0 F 28.3
3:00 272.1 F 28.9
4:00 272.2 F 28.6

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 218.4 F 0.0
10:00 212.5 F 0.0
15:00 225.4 F 0.0
20:00 237.3 F 0.0
25:00 230.9 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:09:37
CYCLE COMPLETE

04/25/2024 09:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____