

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1816

Name: RITTER2  
Sterilizer ID: V2313228  
12/07/2022 11:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.0 F 0.0  
2:00 123.1 F 0.0  
4:00 142.2 F 0.0  
6:00 178.9 F 0.0  
8:00 203.7 F 0.7  
10:00 213.8 F 5.0  
12:00 227.1 F 9.1  
14:00 236.2 F 12.1  
16:00 237.5 F 10.6  
18:00 235.4 F 8.3  
20:00 241.2 F 10.7  
22:00 248.0 F 14.4  
24:00 260.1 F 21.1  
26:00 270.9 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.2 F 29.3  
2:00 272.2 F 28.8  
3:00 271.7 F 28.9  
4:00 272.7 F 29.3

Min 271.5 F 28.4  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 217.4 F 0.0  
10:00 226.6 F 0.0  
15:00 234.6 F 0.0  
20:00 231.7 F 0.0  
25:00 218.6 F 0.0  
30:00 217.9 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 26:38  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:04:36  
CYCLE COMPLETE

12/07/2022 12:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_