
Midmark M11 - v1.0.0
Total Cycles: 2981

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2023 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.9 F 0.0
2:00 181.3 F 4.3
4:00 198.4 F 9.1
6:00 216.4 F 7.4
8:00 231.0 F 9.3
10:00 235.7 F 9.4
12:00 244.0 F 12.6
14:00 242.5 F 11.4
16:00 256.2 F 18.4
18:00 267.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.1 F 28.6
2:00 272.0 F 28.8
3:00 272.0 F 28.4
4:00 272.1 F 28.9

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 230.2 F 0.0
10:00 233.7 F 0.0
15:00 226.1 F 0.0
20:00 227.2 F 0.0
25:00 229.8 F 0.0
30:00 228.9 F 0.0

Drying Complete

FILLING: 1:46
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

03/08/2023 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____