
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003094
12:44 07/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:31
STERILIZING 4:36
136.4 C 221kPa 4:36
CYCLE ABORTED 4:54
STOP BUTTON PRESSED
NOT STERILE
END TIME 4:54

Digital Signature #
285751959B7557BB
