
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001418
11:40 19/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:26
STERILIZING 3:33
136.0 C 220kPa 3:33
136.3 C 224kPa 4:03
136.4 C 224kPa 4:33
135.9 C 221kPa 5:03
136.3 C 224kPa 5:33
136.2 C 223kPa 6:03
136.3 C 224kPa 6:33
136.2 C 223kPa 7:03
136.2 C 223kPa 7:03
VENTING 7:03
AIR DRYING 7:53
CYCLE COMPLETE 22:54

Digital Signature #
4E4C4BA2EB51AF78
