

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5361

Name: RITTER 1  
Sterilizer ID: V2178198  
02/24/2025 04:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.8 F 0.0  
2:00 156.3 F 3.1  
4:00 167.1 F 6.8  
6:00 191.0 F 11.0  
8:00 215.8 F 4.0  
10:00 225.0 F 9.0  
12:00 236.5 F 13.5  
14:00 237.3 F 9.8  
16:00 237.3 F 9.2  
18:00 246.3 F 13.4  
20:00 257.7 F 19.5  
22:00 268.0 F 25.9  
24:00 271.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.5 F 29.1  
2:00 271.9 F 29.3  
3:00 272.2 F 29.0  
4:00 272.1 F 29.4

Min 271.3 F 28.7  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 239.9 F 0.0  
10:00 227.6 F 0.0  
15:00 222.6 F 0.0  
20:00 223.1 F 0.0  
25:00 227.0 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 24:07  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:03:08  
CYCLE COMPLETE

02/24/2025 05:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_