

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1305

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2022 10:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.7 F 0.0  
2:00 191.9 F 2.6  
4:00 210.3 F 7.7  
6:00 226.3 F 12.1  
8:00 231.5 F 9.9  
10:00 235.4 F 8.8  
12:00 243.1 F 12.2  
14:00 245.3 F 13.3  
16:00 257.6 F 20.0  
18:00 268.5 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.3 F 29.4  
2:00 272.1 F 29.8  
3:00 272.1 F 29.2  
4:00 272.3 F 29.9  
  
Min 271.5 F 29.0  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 232.2 F 0.0  
10:00 227.7 F 0.0  
15:00 227.8 F 0.0  
20:00 225.4 F 0.0  
25:00 225.5 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 2:16  
HEATING: 19:10  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

07/15/2022 11:41 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_