
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000372
13:19 31/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:11
STERILIZING 4:13
136.0 C 220kPa 4:13
136.3 C 223kPa 4:43
136.0 C 222kPa 5:13
136.3 C 223kPa 5:43
136.3 C 224kPa 6:13
136.2 C 222kPa 6:43
136.2 C 223kPa 7:13
136.4 C 223kPa 7:43
136.4 C 223kPa 7:43
VENTING 7:43
AIR DRYING 8:39
CYCLE COMPLETE 9:37

Digital Signature #
C2667B5CE33A1534
