
Midmark M11 - v1.0.5
Total Cycles: 473

Name: RITTER2
Sterilizer ID: V2313228
10/22/2021 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 169.8 F 2.9
4:00 181.5 F 7.0
6:00 208.9 F 12.3
8:00 223.1 F 7.4
10:00 236.1 F 12.9
12:00 237.4 F 10.6
14:00 235.8 F 9.2
16:00 250.4 F 16.1
18:00 262.2 F 22.9
20:00 272.9 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.6 F 29.9
2:00 271.9 F 29.2
3:00 271.5 F 29.5
4:00 272.5 F 29.9

Min 271.1 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 218.4 F 0.0
10:00 221.1 F 0.0
15:00 217.1 F 0.0
20:00 218.7 F 0.0
25:00 219.3 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

10/22/2021 11:11 AM
ACCEPTED: X
REJECTED:
Operator: U02