

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000800  
10:00 17/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:13  
PRESSURIZING 6:18  
STERILIZING 7:48  
136.0 C 220kPa 7:48  
136.4 C 223kPa 8:18  
136.0 C 219kPa 8:48  
136.2 C 221kPa 9:18  
136.4 C 223kPa 9:48  
136.1 C 219kPa 10:18  
136.4 C 223kPa 10:48  
135.9 C 219kPa 11:18  
135.9 C 219kPa 11:18  
VENTING 11:18  
AIR DRYING 12:22  
CYCLE COMPLETE 20:48  
-----

Digital Signature #  
C8D057C5BAFCE87B  
-----