

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3500

Name: RITTER 1  
Sterilizer ID: V2178198  
07/07/2023 09:42 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 77.0 F 0.0  
4:00 124.4 F 0.0  
6:00 121.1 F 0.0  
8:00 151.3 F 0.7  
10:00 187.2 F 3.7  
12:00 207.3 F 7.2  
14:00 224.7 F 11.0  
16:00 227.2 F 5.8  
18:00 230.7 F 8.7  
20:00 240.3 F 12.9  
22:00 238.1 F 9.7  
24:00 238.2 F 8.9  
26:00 248.1 F 14.1  
28:00 260.1 F 20.6  
30:00 270.8 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 271.5 F 28.0  
2:00 271.6 F 28.1  
3:00 271.6 F 28.1  
  
Min 271.5 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 234.4 F 0.0  
5:00 214.2 F 0.0  
Dry Cycle Aborted

FILLING: 1:02  
HEATING: 30:39  
STERILIZING: 3:00  
VENTING: 2:49  
DRYING: 5:54  
TOTAL CYCLE: 00:43:24

07/07/2023 10:25 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_