

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002265  
8:22 23/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:46  
STERILIZING 3:40  
136.4 C 220kPa 3:40  
135.9 C 220kPa 4:10  
136.1 C 222kPa 4:40  
136.4 C 224kPa 5:10  
136.4 C 224kPa 5:40  
136.4 C 223kPa 6:10  
136.3 C 221kPa 6:40  
136.3 C 223kPa 7:10  
136.3 C 223kPa 7:10  
VENTING 7:10  
AIR DRYING 8:00  
CYCLE COMPLETE 8:06

-----  
Digital Signature #  
B9DC5019A47AC874  
-----