
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001022
15:11 06/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:28
STERILIZING 3:35
136.0 C 222kPa 3:35
136.2 C 223kPa 4:05
136.4 C 225kPa 4:35
136.4 C 224kPa 5:05
136.4 C 226kPa 5:35
136.4 C 225kPa 6:05
136.4 C 222kPa 6:35
136.4 C 225kPa 7:05
136.4 C 225kPa 7:05
VENTING 7:05
AIR DRYING 7:57
CYCLE COMPLETE 18:50

Digital Signature #
68160867BC4F6B6B
