
Midmark M11 - v1.0.5
Total Cycles: 488

Name: RITTER2
Sterilizer ID: V2313228
10/27/2021 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.8 F 0.0
2:00 104.1 F 0.0
4:00 115.8 F 0.1
6:00 146.1 F 2.4
8:00 174.5 F 5.8
10:00 196.4 F 9.4
12:00 206.8 F 10.0
14:00 225.6 F 8.7
16:00 234.2 F 13.5
18:00 238.8 F 10.7
20:00 234.4 F 8.7
22:00 248.6 F 15.0
24:00 259.5 F 21.2
26:00 269.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 271.8 F 29.7
2:00 272.2 F 29.4
3:00 271.2 F 29.3
4:00 272.5 F 29.8

Min 271.0 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 235.1 F 0.0
10:00 223.5 F 0.0
15:00 222.6 F 0.0
20:00 180.7 F 0.0
25:00 177.5 F 0.0
30:00 180.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:50
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:05:03
CYCLE COMPLETE

10/27/2021 09:44 AM
ACCEPTED: X
REJECTED:
Operator: U01