
Midmark M11 - v1.0.5
Total Cycles: 3208

Name: RITTER2
Sterilizer ID: V2313228
12/12/2023 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.5 F 0.0
2:00 107.0 F 0.1
4:00 110.9 F 1.2
6:00 136.8 F 3.8
8:00 166.6 F 6.7
10:00 182.2 F 10.4
12:00 216.8 F 6.7
14:00 214.9 F 8.0
16:00 227.2 F 12.4
18:00 230.5 F 8.8
20:00 239.8 F 11.7
22:00 242.4 F 12.5
24:00 254.1 F 18.5
26:00 264.9 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 272.5 F 30.1
2:00 271.3 F 29.5
3:00 272.3 F 30.2
4:00 271.7 F 29.6

Min 271.2 F 29.4
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 217.1 F 0.0
10:00 240.2 F 0.0
15:00 230.4 F 0.0
20:00 227.1 F 0.0
25:00 227.0 F 0.0
30:00 225.5 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 27:60
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:06:26
CYCLE COMPLETE

12/12/2023 02:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____