
Midmark M11 - v1.0.0
Total Cycles: 1877

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2022 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.2 F 0.0
2:00 181.8 F 2.1
4:00 186.4 F 6.1
6:00 204.9 F 11.4
8:00 219.9 F 5.3
10:00 232.8 F 10.4
12:00 234.4 F 7.3
14:00 244.9 F 13.4
16:00 243.6 F 11.9
18:00 256.5 F 18.7
20:00 267.8 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.1 F 28.5
2:00 272.5 F 29.0
3:00 271.6 F 28.5
4:00 272.2 F 28.6

Min 271.5 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 226.2 F 0.0
10:00 209.9 F 0.0
15:00 210.1 F 0.0
20:00 226.8 F 0.0
25:00 235.1 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

04/27/2022 09:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____