

-----  
Midmark M11 - v1.0.5  
Total Cycles: 780

Name: RITTER2  
Sterilizer ID: V2313228  
02/10/2022 11:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.6 F 0.0  
2:00 171.6 F 3.7  
4:00 194.4 F 8.5  
6:00 212.1 F 9.7  
8:00 226.3 F 8.7  
10:00 237.2 F 10.9  
12:00 243.2 F 13.0  
14:00 243.8 F 12.7  
16:00 257.1 F 19.8  
18:00 268.9 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.0  
1:00 272.6 F 29.8  
2:00 271.7 F 28.9  
3:00 272.6 F 29.6  
4:00 271.7 F 28.9  
  
Min 271.6 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.7 F 0.0  
5:00 217.3 F 0.0  
10:00 222.7 F 0.0  
15:00 190.8 F 0.0  
20:00 181.3 F 0.0  
25:00 169.0 F 0.0  
30:00 176.9 F 0.0  
  
Drying Complete

FILLING: 2:16  
HEATING: 18:57  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:26  
CYCLE COMPLETE

02/10/2022 12:12 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02