
Midmark M11 - v1.0.5
Total Cycles: 485

Name: RITTER2
Sterilizer ID: V2313228
10/26/2021 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.6 F 0.0
2:00 109.0 F 0.1
4:00 115.8 F 1.7
6:00 143.7 F 5.1
8:00 172.7 F 9.8
10:00 209.7 F 6.5
12:00 229.7 F 9.6
14:00 232.0 F 9.4
16:00 240.9 F 12.4
18:00 240.2 F 11.1
20:00 252.8 F 17.5
22:00 264.5 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.4 F 30.3
2:00 272.8 F 30.4
3:00 272.5 F 30.5
4:00 272.6 F 30.2

Min 270.3 F 29.0
Max 273.7 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 240.6 F 0.0
10:00 227.2 F 0.0
15:00 228.4 F 0.0
20:00 230.1 F 0.0
25:00 193.5 F 0.0
30:00 182.6 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:48
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

10/26/2021 02:08 PM
ACCEPTED: X
REJECTED:
Operator: U01