
Midmark M11 - v1.0.0
Total Cycles: 5004

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2024 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.9 F 0.0
2:00 168.4 F 0.3
4:00 171.4 F 1.1
6:00 193.1 F 5.1
8:00 210.0 F 9.2
10:00 223.1 F 13.5
12:00 226.9 F 8.0
14:00 238.7 F 12.9
16:00 237.8 F 10.2
18:00 236.7 F 8.5
20:00 249.3 F 14.6
22:00 260.4 F 20.8
24:00 270.8 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.8
2:00 272.1 F 28.4
3:00 271.2 F 27.9
4:00 271.7 F 28.6

Min 271.1 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 202.8 F 0.0
10:00 194.8 F 0.0
15:00 195.4 F 0.0
20:00 212.9 F 0.0
25:00 226.9 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:51
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

10/17/2024 10:55 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____