
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000184
10:38 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:11
STERILIZING 4:41
136.0 C 222kPa 4:41
136.3 C 223kPa 5:11
136.4 C 224kPa 5:41
136.4 C 226kPa 6:11
136.2 C 225kPa 6:41
136.3 C 225kPa 7:11
136.2 C 225kPa 7:41
136.4 C 226kPa 8:11
136.4 C 226kPa 8:11
VENTING 8:11
AIR DRYING 9:13
CYCLE COMPLETE 24:13

Digital Signature #
1437D2826533845C
