

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2185

Name: RITTER2  
Sterilizer ID: V2313228  
03/28/2023 10:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.7 F 0.0  
2:00 139.5 F 2.1  
4:00 152.2 F 4.9  
6:00 180.3 F 8.7  
8:00 200.1 F 13.2  
10:00 217.8 F 6.5  
12:00 228.0 F 11.2  
14:00 230.2 F 7.6  
16:00 243.4 F 13.3  
18:00 242.7 F 12.0  
20:00 254.9 F 18.3  
22:00 265.9 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.8  
1:00 272.6 F 30.0  
2:00 272.1 F 29.3  
3:00 272.5 F 30.1  
4:00 272.4 F 29.5

Min 271.0 F 28.7  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.4 F 0.0  
5:00 227.0 F 0.0  
10:00 203.8 F 0.0  
15:00 188.3 F 0.0  
20:00 184.4 F 0.0  
25:00 182.1 F 0.0  
30:00 181.6 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 23:45  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:02:07  
CYCLE COMPLETE

03/28/2023 12:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_