
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001555
8:52 11/04/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:00
CASSETTE REMOVED
NOT STERILE
END TIME 0:00

Digital Signature #
80A86B324BA5F8DF
