
Midmark M11 - v1.0.5
Total Cycles: 3191

Name: RITTER2
Sterilizer ID: V2313228
12/06/2023 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.2 F 0.0
2:00 125.9 F 0.8
4:00 150.6 F 3.4
6:00 174.1 F 7.2
8:00 197.0 F 11.3
10:00 213.8 F 5.0
12:00 224.1 F 9.8
14:00 231.6 F 9.1
16:00 239.8 F 11.8
18:00 239.1 F 10.6
20:00 252.2 F 17.1
22:00 263.7 F 23.8
24:00 271.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.6 F 29.8
2:00 271.4 F 29.1
3:00 271.8 F 29.8
4:00 272.4 F 29.7

Min 270.9 F 28.9
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 210.0 F 0.0
10:00 196.3 F 0.0
15:00 195.4 F 0.0
20:00 194.5 F 0.0
25:00 196.1 F 0.0
30:00 197.3 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:41
CYCLE COMPLETE

12/06/2023 02:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____