

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002202  
8:47 04/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:31  
PRESSURIZING 3:33  
STERILIZING 4:44  
136.4 C 223kPa 4:44  
136.4 C 225kPa 5:14  
136.4 C 223kPa 5:44  
136.4 C 224kPa 6:14  
136.4 C 223kPa 6:44  
136.4 C 224kPa 7:14  
136.2 C 221kPa 7:44  
136.4 C 223kPa 8:14  
136.4 C 223kPa 8:14  
VENTING 8:14  
AIR DRYING 9:06  
CYCLE COMPLETE 39:06  
-----

Digital Signature #  
6271CF7302173A91  
-----