
Midmark M11 - v1.0.0
Total Cycles: 1866

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2022 04:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.4 F 0.0
2:00 165.7 F 2.8
4:00 180.5 F 6.9
6:00 206.2 F 12.5
8:00 221.4 F 6.3
10:00 235.1 F 11.3
12:00 233.5 F 8.1
14:00 242.1 F 10.8
16:00 245.9 F 13.1
18:00 258.6 F 20.0
20:00 269.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 28.8
2:00 272.0 F 28.9
3:00 272.0 F 28.6
4:00 272.2 F 29.2

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 205.8 F 0.0
10:00 209.6 F 0.0
15:00 207.5 F 0.0
20:00 228.0 F 0.0
25:00 232.8 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

04/25/2022 05:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____