
Midmark M11 - v1.0.5
Total Cycles: 1456

Name: RITTER2
Sterilizer ID: V2313228
09/06/2022 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 190.2 F 2.6
4:00 197.3 F 7.7
6:00 216.9 F 12.1
8:00 232.7 F 9.2
10:00 237.2 F 9.7
12:00 244.5 F 13.4
14:00 244.6 F 12.8
16:00 257.6 F 19.9
18:00 269.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.4 F 29.1
2:00 272.7 F 29.5
3:00 272.0 F 29.3
4:00 272.2 F 29.1

Min 271.8 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 222.6 F 0.0
10:00 220.4 F 0.0
15:00 245.1 F 0.0
20:00 226.3 F 0.0
25:00 226.1 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 19:00
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

09/06/2022 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____