

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1202

Name: RITTER2  
Sterilizer ID: V2313228  
06/20/2022 12:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.8 F 0.0  
2:00 124.7 F 1.5  
4:00 133.6 F 4.1  
6:00 159.9 F 7.9  
8:00 193.1 F 12.5  
10:00 217.5 F 6.1  
12:00 231.8 F 10.9  
14:00 230.7 F 7.1  
16:00 241.5 F 12.5  
18:00 240.0 F 10.6  
20:00 251.8 F 16.7  
22:00 262.4 F 22.9  
24:00 272.6 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.2 F 29.6  
2:00 271.2 F 28.9  
3:00 272.4 F 29.8  
4:00 271.6 F 29.3

Min 271.1 F 28.9  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
5:00 208.5 F 0.0  
10:00 227.3 F 0.0  
15:00 228.9 F 0.0  
20:00 230.4 F 0.0  
25:00 226.1 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 24:27  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:02:37  
CYCLE COMPLETE

06/21/2022 06:46 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02