
Midmark M11 - v1.0.5
Total Cycles: 3475

Name: RITTER2
Sterilizer ID: V2313228
03/25/2024 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.6 F 0.0
2:00 110.4 F 0.0
4:00 114.5 F 0.1
6:00 147.5 F 1.6
8:00 167.8 F 4.9
10:00 190.0 F 8.5
12:00 208.3 F 12.5
14:00 216.1 F 7.2
16:00 229.0 F 12.0
18:00 235.9 F 9.3
20:00 239.2 F 10.1
22:00 247.1 F 13.9
24:00 257.7 F 20.0
26:00 267.6 F 26.6
28:00 271.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 273.0 F 30.0
2:00 271.3 F 29.3
3:00 273.0 F 30.1
4:00 271.9 F 29.6

Min 271.2 F 29.3
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 240.0 F 0.0
10:00 229.2 F 0.0
15:00 210.7 F 0.0
20:00 204.0 F 0.0
25:00 214.3 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:05
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:06:35
CYCLE COMPLETE

03/25/2024 02:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____