
Midmark M11 - v1.0.0
Total Cycles: 1559

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2022 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.0 F 0.0
2:00 192.1 F 3.4
4:00 181.0 F 7.7
6:00 199.1 F 12.5
8:00 223.8 F 6.5
10:00 235.0 F 11.6
12:00 235.5 F 8.2
14:00 243.4 F 11.3
16:00 245.2 F 12.5
18:00 257.0 F 18.8
20:00 267.5 F 25.2
22:00 272.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 271.9 F 28.4
2:00 272.6 F 28.7
3:00 272.2 F 28.9
4:00 271.3 F 28.3

Min 271.2 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 229.1 F 0.0
10:00 228.3 F 0.0
15:00 230.6 F 0.0
20:00 229.1 F 0.0
25:00 229.2 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

01/25/2022 12:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____