
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001488
12:03 26/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:08
STERILIZING 4:11
136.0 C 222kPa 4:11
136.4 C 223kPa 4:41
136.3 C 225kPa 5:11
136.2 C 224kPa 5:41
136.3 C 225kPa 6:11
136.4 C 226kPa 6:41
136.4 C 226kPa 7:11
136.1 C 224kPa 7:41
136.1 C 224kPa 7:41
VENTING 7:41
AIR DRYING 8:37
CYCLE COMPLETE 8:43

Digital Signature #
3561582B63DD3500
