
Midmark M11 - v1.0.5
Total Cycles: 3643

Name: RITTER2
Sterilizer ID: V2313228
05/24/2024 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 149.2 F 1.7
4:00 162.9 F 4.7
6:00 186.9 F 8.9
8:00 202.0 F 11.7
10:00 220.0 F 7.8
12:00 232.2 F 12.6
14:00 233.3 F 9.6
16:00 237.6 F 9.6
18:00 246.2 F 14.2
20:00 257.9 F 20.6
22:00 268.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 273.1 F 30.2
2:00 271.7 F 29.5
3:00 271.4 F 29.8
4:00 273.1 F 30.1

Min 271.1 F 29.3
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 210.6 F 0.0
10:00 210.4 F 0.0
15:00 199.4 F 0.0
20:00 196.0 F 0.0
25:00 198.0 F 0.0
30:00 200.8 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 23:19
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

05/24/2024 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____