

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5335

Name: RITTER 1  
Sterilizer ID: V2178198  
02/12/2025 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.7 F 0.0  
2:00 75.0 F 0.4  
4:00 90.7 F 2.1  
6:00 134.1 F 4.6  
8:00 158.0 F 7.6  
10:00 173.5 F 11.1  
12:00 186.9 F 2.3  
14:00 207.0 F 4.9  
16:00 218.2 F 8.3  
18:00 229.2 F 12.2  
20:00 229.6 F 7.2  
22:00 241.6 F 12.3  
24:00 239.2 F 9.9  
26:00 251.6 F 16.0  
28:00 262.2 F 22.1  
30:00 272.1 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.8 F 28.9  
2:00 271.9 F 28.1  
3:00 272.6 F 29.0  
4:00 272.6 F 28.6

Min 271.6 F 28.1  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
5:00 185.9 F 0.0  
10:00 198.2 F 0.0  
15:00 200.4 F 0.0  
20:00 184.9 F 0.0  
25:00 183.8 F 0.0  
30:00 187.7 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 30:34  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:08:46  
CYCLE COMPLETE

02/12/2025 10:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_