
Midmark M11 - v1.0.5
Total Cycles: 4455

Name: RITTER2
Sterilizer ID: V2313228
03/25/2025 03:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.0 F 0.0
2:00 145.5 F 1.8
4:00 159.6 F 5.4
6:00 200.2 F 9.4
8:00 223.1 F 7.6
10:00 228.1 F 7.8
12:00 238.6 F 12.3
14:00 234.7 F 9.2
16:00 239.2 F 10.4
18:00 245.0 F 13.3
20:00 256.3 F 19.5
22:00 267.0 F 26.1
24:00 273.6 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.0
1:00 272.9 F 30.1
2:00 272.6 F 30.6
3:00 271.2 F 29.9
4:00 273.3 F 30.4

Min 271.1 F 29.9
Max 273.4 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 228.0 F 0.0
10:00 232.2 F 0.0
15:00 227.7 F 0.0
20:00 213.3 F 0.0
25:00 225.3 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:12
CYCLE COMPLETE

03/25/2025 04:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____