
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
1.9uS / 1.2ppm
CYCLE NUMBER 000040
16:55 10/08/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:26
PRESSURIZING 2:35
STERILIZING 3:55
136.4 C 227kPa 3:55
136.4 C 227kPa 4:25
136.3 C 226kPa 4:55
136.4 C 226kPa 5:25
136.3 C 226kPa 5:55
136.1 C 224kPa 6:25
136.4 C 227kPa 6:55
136.0 C 225kPa 7:25
VENTING 7:25
AIR DRYING 8:16
CYCLE COMPLETE 10:17

Digital Signature #
F690945D5CD8A18F
