

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5437

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2025 03:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.3 F 0.0  
2:00 127.1 F 1.6  
4:00 154.1 F 4.7  
6:00 165.3 F 8.3  
8:00 189.1 F 12.2  
10:00 212.0 F 3.4  
12:00 217.0 F 7.3  
14:00 229.8 F 11.3  
16:00 230.4 F 6.9  
18:00 240.8 F 12.3  
20:00 239.4 F 10.3  
22:00 251.5 F 16.2  
24:00 262.0 F 22.4  
26:00 271.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.0 F 29.3  
2:00 271.2 F 28.9  
3:00 271.9 F 29.4  
4:00 271.2 F 28.8

Min 271.1 F 28.7  
Max 272.1 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 195.3 F 0.0  
10:00 192.0 F 0.0  
15:00 195.0 F 0.0  
20:00 197.6 F 0.0  
25:00 202.2 F 0.0  
30:00 202.4 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 26:35  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:05:22  
CYCLE COMPLETE

03/24/2025 04:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_