

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5067

Name: RITTER 1  
Sterilizer ID: V2178198  
11/07/2024 12:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.3 F 0.0  
2:00 144.7 F 0.3  
4:00 165.5 F 1.6  
6:00 196.0 F 5.8  
8:00 211.0 F 10.3  
10:00 222.2 F 6.6  
12:00 231.0 F 8.6  
14:00 241.2 F 13.3  
16:00 240.0 F 10.6  
18:00 236.0 F 8.6  
20:00 250.7 F 15.3  
22:00 261.9 F 21.6  
24:00 272.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.6 F 28.8  
2:00 272.0 F 28.4  
3:00 271.3 F 27.9  
4:00 271.8 F 28.5

Min 271.2 F 27.8  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.6 F 0.0  
5:00 212.7 F 0.0  
10:00 210.5 F 0.0  
15:00 208.0 F 0.0  
20:00 201.0 F 0.0  
25:00 205.0 F 0.0  
30:00 204.2 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 24:34  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:03:15  
CYCLE COMPLETE

11/07/2024 01:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_