
Midmark M11 - v1.0.0
Total Cycles: 1040

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2021 01:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.0 F 0.0
2:00 187.2 F 3.6
4:00 206.3 F 8.7
6:00 219.9 F 9.7
8:00 231.5 F 9.5
10:00 234.5 F 9.0
12:00 244.6 F 13.1
14:00 243.6 F 12.1
16:00 256.8 F 18.9
18:00 268.2 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.4 F 28.7
2:00 272.2 F 28.9
3:00 272.2 F 28.5
4:00 272.2 F 28.9

Min 271.6 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 219.3 F 0.0
10:00 224.8 F 0.0
15:00 219.6 F 0.0
20:00 213.5 F 0.0
25:00 216.5 F 0.0
Dry Cycle Aborted

FILLING: 2:15
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 25:56
TOTAL CYCLE: 00:54:31

07/21/2021 02:52 PM
ACCEPTED: X
REJECTED:
Operator: U01