
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001142
7:16 09/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:53
STERILIZING 3:58
136.4 C 223kPa 3:58
136.4 C 224kPa 4:28
136.3 C 223kPa 4:58
136.4 C 224kPa 5:28
136.4 C 223kPa 5:58
136.4 C 219kPa 6:28
136.4 C 219kPa 6:58
136.4 C 221kPa 7:28
136.4 C 221kPa 7:28
VENTING 7:28
AIR DRYING 8:20
CYCLE COMPLETE 8:22

Digital Signature #
947CFB689E5EEE1C
