
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001310
8:14 02/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:26
STERILIZING 4:36
136.0 C 220kPa 4:36
136.4 C 224kPa 5:06
136.4 C 223kPa 5:36
136.0 C 222kPa 6:06
136.4 C 224kPa 6:36
136.4 C 224kPa 7:06
136.1 C 217kPa 7:36
136.4 C 223kPa 8:06
136.4 C 223kPa 8:06
VENTING 8:06
AIR DRYING 9:02
CYCLE COMPLETE 9:06

Digital Signature #
DA9937F363707B98
