

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3173

Name: RITTER2  
Sterilizer ID: V2313228  
11/30/2023 03:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.7 F 0.0  
2:00 94.3 F 0.2  
4:00 109.2 F 1.5  
6:00 144.5 F 4.5  
8:00 169.5 F 8.0  
10:00 194.4 F 12.0  
12:00 211.6 F 5.8  
14:00 224.3 F 10.5  
16:00 228.2 F 7.5  
18:00 238.9 F 12.0  
20:00 237.7 F 10.2  
22:00 250.8 F 16.5  
24:00 262.3 F 23.1  
26:00 272.8 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.7  
1:00 271.8 F 29.6  
2:00 271.9 F 29.6  
3:00 272.0 F 29.6  
4:00 272.1 F 29.8

Min 271.8 F 29.6  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 216.2 F 0.0  
10:00 193.8 F 0.0  
15:00 191.1 F 0.0  
20:00 189.5 F 0.0  
25:00 189.2 F 0.0  
30:00 189.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 27:02  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:05:13  
CYCLE COMPLETE

11/30/2023 04:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_