
Midmark M11 - v1.0.5
Total Cycles: 4108

Name: RITTER2
Sterilizer ID: V2313228
11/12/2024 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.3 F 0.0
2:00 73.4 F 0.0
4:00 113.7 F 0.0
6:00 110.8 F 0.1
8:00 145.7 F 0.4
10:00 174.6 F 0.6
12:00 194.2 F 2.1
14:00 211.3 F 5.2
16:00 220.9 F 8.7
18:00 230.8 F 12.5
20:00 231.4 F 8.3
22:00 242.9 F 13.3
24:00 241.0 F 11.3
26:00 239.0 F 10.0
28:00 251.9 F 16.4
30:00 262.6 F 22.6
32:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 28.9
2:00 272.3 F 29.3
3:00 271.6 F 28.4
4:00 272.6 F 29.1

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 209.5 F 0.0
10:00 206.0 F 0.0
15:00 200.6 F 0.0
20:00 200.8 F 0.0
25:00 201.8 F 0.0
30:00 201.4 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:24
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:10:34
CYCLE COMPLETE

11/12/2024 10:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____