
Midmark M11 - v1.0.5
Total Cycles: 504

Name: RITTER2
Sterilizer ID: V2313228
10/29/2021 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.2 F 0.0
2:00 172.9 F 3.6
4:00 187.6 F 8.5
6:00 210.4 F 9.0
8:00 226.9 F 8.8
10:00 236.9 F 10.6
12:00 242.1 F 12.5
14:00 242.3 F 11.9
16:00 255.8 F 18.9
18:00 267.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.2 F 29.4
2:00 273.0 F 29.7
3:00 272.2 F 29.2
4:00 271.6 F 28.9

Min 271.5 F 28.8
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 227.9 F 0.0
10:00 222.1 F 0.0
15:00 219.9 F 0.0
20:00 225.4 F 0.0
25:00 226.3 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

10/29/2021 12:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____