
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000644
10:43 10/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:48
STERILIZING 3:59
136.0 C 224kPa 3:59
136.2 C 231kPa 4:29
136.4 C 231kPa 4:59
136.4 C 230kPa 5:29
136.3 C 232kPa 5:59
136.4 C 232kPa 6:29
136.4 C 232kPa 6:59
136.3 C 232kPa 7:29
136.3 C 232kPa 7:29
VENTING 7:29
AIR DRYING 8:19
CYCLE COMPLETE 23:19

Digital Signature #
D818401521B081C7
