
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005796
9:29 20/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:04
STERILIZING 4:17
136.4 C 220kPa 4:17
136.2 C 224kPa 4:47
136.4 C 224kPa 5:17
136.4 C 222kPa 5:47
136.1 C 223kPa 6:17
136.2 C 224kPa 6:47
136.4 C 223kPa 7:17
136.4 C 226kPa 7:47
136.4 C 226kPa 7:47
VENTING 7:47
AIR DRYING 9:00
CYCLE COMPLETE 9:04

Digital Signature #
AE494131A7DE4D94
