
Midmark M11 - v1.0.0
Total Cycles: 4748

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2024 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.7 F 0.0
2:00 155.6 F 2.1
4:00 159.2 F 5.2
6:00 182.7 F 8.6
8:00 201.0 F 13.0
10:00 212.6 F 5.1
12:00 225.1 F 9.3
14:00 236.2 F 11.8
16:00 236.3 F 10.0
18:00 237.6 F 9.1
20:00 246.2 F 13.4
22:00 257.2 F 19.3
24:00 267.3 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 271.6 F 28.7
2:00 272.6 F 29.1
3:00 272.0 F 28.6
4:00 271.2 F 28.3

Min 271.2 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 207.0 F 0.0
10:00 215.1 F 0.0
15:00 225.5 F 0.0
20:00 239.8 F 0.0
25:00 236.8 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:25
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:06
CYCLE COMPLETE

07/19/2024 10:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____