
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001289
14:00 11/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:42
STERILIZING 3:52
136.0 C 220kPa 3:52
136.4 C 223kPa 4:23
136.4 C 221kPa 4:53
136.0 C 222kPa 5:23
136.2 C 223kPa 5:53
136.4 C 223kPa 6:23
136.4 C 225kPa 6:53
136.1 C 222kPa 7:23
136.1 C 222kPa 7:23
VENTING 7:23
AIR DRYING 8:13
CYCLE COMPLETE 23:13

Digital Signature #
6E72A506902596F9
