

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1780

Name: RITTER 1  
Sterilizer ID: V2178198  
04/01/2022 07:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	123.1 F	0.0
2:00	131.0 F	0.0
4:00	141.2 F	0.0
6:00	174.9 F	0.0
8:00	198.9 F	0.0
10:00	210.5 F	0.0
12:00	211.6 F	0.0
14:00	211.7 F	0.0
16:00	211.7 F	0.0
18:00	211.7 F	0.0
20:00	211.7 F	0.0
22:00	211.7 F	0.0
24:00	211.7 F	0.0
26:00	211.7 F	0.0
28:00	211.6 F	0.0

C633: HEATUP MODE  
PRESSURE LOW  
CHAMBER TEMP 212F  
CHAMBER PRESS 0.0PSI  
\*\* ITEMS NOT STERILE \*\*

04/01/2022 07:40 AM