
Midmark M11 - v1.0.0
Total Cycles: 1881

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2022 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.4 F 0.0
2:00 147.9 F 0.0
4:00 160.2 F 2.5
6:00 184.5 F 6.0
8:00 204.7 F 10.5
10:00 223.4 F 5.0
12:00 229.3 F 9.3
14:00 238.9 F 10.2
16:00 241.4 F 11.8
18:00 239.6 F 10.1
20:00 253.2 F 16.8
22:00 264.7 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.6 F 29.1
2:00 272.0 F 28.5
3:00 271.4 F 28.3
4:00 272.7 F 29.1

Min 271.2 F 28.0
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 191.9 F 0.0
10:00 212.4 F 0.0
15:00 229.8 F 0.0
20:00 230.4 F 0.0
25:00 229.0 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:58
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

04/27/2022 02:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____