
Midmark M11 - v1.0.0
Total Cycles: 4954

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2024 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.1 F 0.0
2:00 120.8 F 0.3
4:00 131.5 F 1.3
6:00 166.7 F 3.9
8:00 196.0 F 7.2
10:00 217.4 F 11.0
12:00 222.9 F 4.4
14:00 224.3 F 7.7
16:00 235.7 F 11.7
18:00 232.9 F 7.7
20:00 243.1 F 12.9
22:00 241.2 F 10.9
24:00 252.5 F 16.5
26:00 262.4 F 22.3
28:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.2 F 29.0
2:00 272.3 F 28.9
3:00 271.8 F 28.5
4:00 271.2 F 28.3

Min 271.2 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 216.6 F 0.0
10:00 203.9 F 0.0
15:00 204.5 F 0.0
20:00 231.9 F 0.0
25:00 228.6 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:42
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:07:22
CYCLE COMPLETE

10/02/2024 02:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____