
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002667
9:24 08/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:50
STERILIZING 3:58
136.0 C 220kPa 3:58
136.4 C 226kPa 4:28
136.3 C 223kPa 4:58
136.4 C 223kPa 5:28
136.1 C 223kPa 5:58
136.1 C 223kPa 6:28
136.3 C 224kPa 6:58
136.4 C 225kPa 7:28
136.4 C 225kPa 7:28
VENTING 7:28
AIR DRYING 8:20
CYCLE COMPLETE 16:10

Digital Signature #
D3374A2D132C08D9
